#### §870.1025

- (1) For use in the diagnosis, monitoring, or screening of neoplastic diseases with the exception of immunohistochemical devices;
- (2) For use in screening or diagnosis of familial or acquired genetic disorders, including inborn errors of metabolism:
- (3) For measuring an analyte that serves as a surrogate marker for screening, diagnosis, or monitoring life-threatening diseases such as acquired immune deficiency syndrome (AIDS), chronic or active hepatitis, tuberculosis, or myocardial infarction or to monitor therapy;
- (4) For assessing the risk of cardiovascular diseases;
  - (5) For use in diabetes management;
- (6) For identifying or inferring the identity of a microorganism directly from clinical material;
- (7) For detection of antibodies to microorganisms other than immunoglobulin G (IgG) or IgG assays when the results are not qualitative, or are used to determine immunity, or the assay is intended for use in matrices other than serum or plasma;
- (8) For noninvasive testing as defined in  $\S 812.3(k)$  of this chapter; and
- (9) For near patient testing (point of

[65 FR 2314, Jan. 14, 2000]

### Subpart B—Cardiovascular Diagnostic Devices

#### §870.1025 Arrhythmia detector and alarm (including ST-segment measurement and alarm).

- (a) Identification. The arrhythmia detector and alarm device monitors an electrocardiogram and is designed to produce a visible or audible signal or alarm when atrial or ventricular arrhythmia, such as premature contraction or ventricular fibrillation, occurs.
- (b) Classification. Class II (special controls). The guidance document entitled "Class II Special Controls Guidance Document: Arrhythmia Detector and Alarm" will serve as the special control. See §870.1 for the availability of this guidance document.

[68 FR 61344, Oct. 28, 2003]

#### §870.1100 Blood pressure alarm.

- (a) *Identification*. A blood pressure alarm is a device that accepts the signal from a blood pressure transducer amplifier, processes the signal, and emits an alarm when the blood pressure falls outside a pre-set upper or lower limit.
- (b) Classification. Class II (performance standards).

#### §870.1110 Blood pressure computer.

- (a) *Identification*. A blood pressure computer is a device that accepts the electrical signal from a blood pressure transducer amplifier and indicates the systolic, diastolic, or mean pressure based on the input signal.
- (b) Classification. Class II (performance standards).

#### §870.1120 Blood pressure cuff.

- (a) *Identification*. A blood pressure cuff is a device that has an inflatable bladder in an inelastic sleeve (cuff) with a mechanism for inflating and deflating the bladder. The cuff is used in conjunction with another device to determine a subject's blood pressure.
- (b) Classification. Class II (performance standards).

# §870.1130 Noninvasive blood pressure measurement system.

- (a) *Identification*. A noninvasive blood pressure measurement system is a device that provides a signal from which systolic, diastolic, mean, or any combination of the three pressures can be derived through the use of tranducers placed on the surface of the body.
- (b) Classification. Class II (performance standards).

### §870.1140 Venous blood pressure manometer.

- (a) *Identification*. A venous blood pressure manometer is a device attached to a venous catheter to indicate manometrically the central or peripheral venous pressure.
- (b) Classification. Class II (performance standards).

### § 870.1200 Diagnostic intravascular catheter.

(a) Identification. An intravascular diagnostic catheter is a device used to

### Food and Drug Administration, HHS

record intracardiac pressures, to sample blood, and to introduce substances into the heart and vessels. Included in this generic device are right-heart catheters, left-heart catheters, and angiographic catheters, among others.

(b) Classification. Class II (performance standards).

#### §870.1210 Continuous flush catheter.

- (a) *Identification*. A continuous flush catheter is an attachment to a catheter-transducer system that permits continuous intravascular flushing at a slow infusion rate for the purpose of eliminating clotting, back-leakage, and waveform damping.
- (b) Classification. Class II (performance standards).

## § 870.1220 Electrode recording catheter or electrode recording probe.

- (a) Identification. An electrode recording catheter or an electrode recording probe is a device used to detect an intracardiac electrocardiogram, or to detect cardiac output or left-to-right heart shunts. The device may be unipolar or multipolar for electrocardiogram detection, or may be a platinum-tipped catheter which senses the presence of a special indicator for cardiac output or left-to-right heart shunt determinations.
- (b) Classification. Class II (performance standards).

# \$870.1230 Fiberoptic oximeter catheter.

- (a) Identification. A fiberoptic oximeter catheter is a device used to estimate the oxygen saturation of the blood. It consists of two fiberoptic bundles that conduct light at a desired wavelength through blood and detect the reflected and scattered light at the distal end of the catheter.
- (b) Classification. Class II (performance standards).

#### §870.1240 Flow-directed catheter.

- (a) *Identification*. A flow-directed catheter is a device that incorporates a gas-filled balloon to help direct the catheter to the desired position.
- (b) Classification. Class II (performance standards).

#### §870.1250 Percutaneous catheter.

- (a) *Identification*. A percutaneous catheter is a device that is introduced into a vein or artery through the skin using a dilator and a sheath (introducer) or guide wire.
- (b) Classification. Class II (performance standards).

## § 870.1270 Intracavitary phonocatheter system.

- (a) Identification. An intracavitary phonocatheter system is a system that includes a catheter with an acoustic transducer and the associated device that processes the signal from the transducer; this device records bioacoustic phenomena from a transducer placed within the heart, blood vessels, or body cavities.
- (b) Classification. Class II (performance standards).

#### § 870.1280 Steerable catheter.

- (a) *Identification*. A steerable catheter is a catheter used for diagnostic and monitoring purposes whose movements are directed by a steering control unit.
- (b) Classification. Class II (performance standards).

### §870.1290 Steerable catheter control system.

- (a) *Identification*. A steerable catheter control system is a device that is connected to the proximal end of a steerable guide wire that controls the motion of the steerable catheter.
- (b)  ${\it Classification}$ . Class II (performance standards).

#### §870.1300 Catheter cannula.

- (a) *Identification*. A catheter cannula is a hollow tube which is inserted into a vessel or cavity; this device provides a rigid or semirigid structure which can be connected to a tube or connector.
- (b) Classification. Class II (performance standards).

# § 870.1310 Vessel dilator for percutaneous catheterization.

(a) *Identification*. A vessel dilator for percutaneous catheterization is a device which is placed over the guide wire to enlarge the opening in the vessel, and which is then removed before sliding the catheter over the guide wire.